



## **Lutetium (Lu-177)**

---

**Half Life:** 6.7 days

**Radiation:** Beta and a low energy gamma

**Shielding:** 2.1 mm lead for gamma, 1.2 mm plexiglass for beta

**Dosimetry:** Body and ring badges. Urine assays may be required after spills or contamination incidents

**Detection/Measurement:** GM meter with pancake probe. Wipe tests

### **General Precautions:**

- **Maintain your occupational exposure to radiation As Low As Reasonably Achievable [ALARA]**
- **All persons handling radioactive material must be properly trained by EHS prior to handling and are listed as a rad worker by EHS department**
- **Plan experiments accordingly to minimize external exposure by reducing exposure time, using shielding and increasing your distance from the radiation source**
- **Monitor yourself and the work area during and after each use of radioactive material**
- **Use the smallest amount of radioisotope possible to minimize radiation dose and radioactive waste**
- **Keep an accurate inventory of all radioactive material including records of all receipts, transfers and disposal – contact EHS for any disposal needs including liquid waste**
- **Perform and record lab surveys as needed (monthly and post experiment)**
- **Avoid generating mixed waste (combinations of radioactive, biological and chemical waste)**

### **Special Precautions:**

- Store behind 2.1 mm lead shielding when working with Lutetium 177
- Use tools to indirectly handle unshielded sources and potentially contaminated containers - no direct hand contact
- Ensure that an appropriate, operational survey meter is present in the work area and turned on whenever lutetium is handled to immediately detect contamination
- Shield waste containers as needed to maintain accessible dose rate ALARA

### **Safe Lab Practices:**

- Disposable gloves, lab coats, and safety glasses are the minimum PPE required when handling radioactive material
- Remove and discard potentially contaminated PPE prior to leaving the lab area where radioactive material is used
- Cover all lab bench tops where radioactive material is handled with plastic-backed absorbent paper – change as needed
- Handle radioactive solutions in trays large enough to contain the material in the event of a spill
- Never eat, drink, smoke, handle contact lenses, apply cosmetics, or take medicine in the lab - keep food, drinks, and cosmetics out of the lab entirely
- Never pipette by mouth
- Never store food and beverages in refrigerators/freezers used for storing radioisotopes – ensure that isotopes are secured at all times
- Avoid any skin contact with skin-absorbable solvents containing radioactive materials
- Fume hoods and biological safety cabinets for use with non-airborne radioactive material must work properly and be inspected annually by the EHS department
- Do not take any radioactive material off site or use in any ways not approved by the Radiation Safety Officer