


Case Center for Imaging Research

Survey Report – Radionuclide Use Area

Location: Optical Authorized R. Muzic Radionuclides Used: _____
Lerner B210 User: _____

Post experiment Weekly Semiannual

Survey map



Trigger 200 dpm
Level: /100 cm²

METER SURVEY:

Instrument*
Make: Ludlum
Model: 14C
Serial #: 269852

Calibration
Date: 7/11/2014
Efficiency: 0.82%
Op. Check: _____

Results:

No readings above trigger level Other: _____

WIPE TEST RESULTS:

Trigger 200 dpm
Level: /100 cm²

Instrument: Wallac LKB 1282 (SN 446) Efficiency: 82.4%

See reverse page for counts

SUMMARY OF FINDINGS:

No significant contamination levels / meter readings
 Other: _____

Surveyed by (print & sign)

Date

* Meter calibration sheet on file.

Case Center for Imaging Research

Survey Report – Radionuclide Use Area

Date : _____

Name : _____

Location	Survey			Wipe Test			Notes
	cpm	dpm	< 2x Bkg?	cpm	dpm	< 2x Bkg?	dpm = (cpm – cpm bkgr)/efficiency
EFFICIENCY							
Background							
1 – bench							
2 – counter							
3 – floor							
4 – floor							

* Values required only when cpm exceed 2x background.