

Case Center for Imaging Research

Survey Report – Radionuclide Use Area

Location: **Surgery suite 1
Lerner B221**

Authorized User: **R. Muzic**


Radionuclides Used: _____

Post experiment

Weekly

Semiannual

Survey map



METER SURVEY:

Trigger 200 dpm
Level: /100 cm²

Instrument*

Make: Ludlum
Model: 14C
Serial #: 269852

Calibration

Date: 7/11/2014
Efficiency: 0.82%
Op. Check: _____

Results:

No readings above trigger level

Other: _____

WIPE TEST RESULTS:

Trigger 200 dpm
Level: /100 cm²

Instrument: Wallac LKB 1282 (SN 446)

Efficiency: 82.4%

See reverse page for counts

SUMMARY OF FINDINGS:

No significant contamination levels / meter readings

Other:
: _____

Surveyed by (print & sign)

Date

* Meter calibration sheet on file.

Case Center for Imaging Research

Survey Report – Radionuclide Use Area

Date : _____

Name : _____

Location	Survey			Wipe Test			Notes
	cpm	dpm	< 2x Bkg?	cpm	dpm	< 2x Bkg?	
EFFICIENCY							dpm = (cpm – cpm bkgr)/efficiency
Background							
1 – counter							
2 – floor							
3 – hood							
4 – floor							

* Values required only when cpm exceed 2x background.