


Case Center for Imaging Research

Survey Report – Radionuclide Use Area

Location: **Micro PET-MRI Lerner B207B** Authorized **Muzic** Radionuclides Used: _____
User: _____

Post Experiment Weekly Semiannual

Survey map



Trigger 200 dpm
Level: /100 cm²

METER SURVEY:

Instrument* Ludlum
Make: _____
Model: 14C
Serial #: 269852

Calibration
Date: 07/27/2021
Efficiency: 0.96%
Op. Check: _____

Results:

No readings above trigger level Other: _____

WIPE TEST RESULTS:

Trigger 200 dpm
Level: /100 cm²

Instrument: Hidex AMG (SN 2180119) Efficiency: 46% (11-19-20)

See reverse page for counts

SUMMARY OF FINDINGS:

No significant contamination levels / meter readings
 Other: _____

Surveyed by (print & sign)

Date

* Meter calibration sheet on file.

Case Center for Imaging Research

Survey Report – Radionuclide Use Area

Date : _____

Name : _____

Location	Survey			Wipe Test			Notes
	cpm	dpm	< 2x Bkg?	cpm	dpm	< 2x Bkg?	dpm = (cpm – cpm bkgr)/efficiency
EFFICIENCY							
Background							
1 – bench							
2 – scanner							
3 – table							
4 – floor							

* Values required only when cpm exceed 2x background.